

**For more detailed information on the standards mentioned in the PowerPoint presentation ‘Air kerma and absorbed dose standards for reference dosimetry in brachytherapy’, T Sander, (International Workshop on Metrology for Brachytherapy, PTB, Braunschweig, Germany, 30 – 31 May 2017), please refer to the following references:**

#### **[ARPANSA \(Australia\)](#)**

Poder J, Smith R L, Shelton N, Whitaker M, Butler D and Haworth A. High dose rate brachytherapy source measurement intercomparison. *Australas Phys Eng Sci Med* 2017; doi:10.1007/s13246-017-0542-6

Butler D, Haworth A, Sander T and Todd S. Comparison of <sup>192</sup>Ir air kerma calibration coefficients derived at ARPANSA using the interpolation method and at the National Physical Laboratory using a direct measurement. *Australas Phys Eng Sci Med* 2008; 31: 332–8.

#### **[ENEA-INMRI \(Italy\)](#)**

Toni M P, Pimpinella M, Pinto M, Quini M, Cappadozzi G, Silvestri C and Bottauscio O. Direct determination of the absorbed dose to water from <sup>125</sup>I low dose-rate brachytherapy seeds using the new absorbed dose primary standard developed at ENEA-INMRI. *Metrologia* 2012; 49: S193–7.

Guerra A S, Loreti S, Pimpinella M, Quini M, D’Arienzo M, Astefanoaei I, Caporali C, Bolzan C and Pagliari M. A standard graphite calorimeter for dosimetry in brachytherapy with high dose rate <sup>192</sup>Ir sources. *Metrologia* 2012; 49: S179–83.

Piermattei A and Azario L. Applications of the Italian protocol for the calibration of brachytherapy sources. *Phys Med Biol* 1997; 42 1661–9.

#### **[NMIJ-AIST \(Japan\)](#)**

Kessler C, Kurosawa T and Mikamoto T. Comparison BIPM.RI(I)-K8 of high dose-rate Ir-192 brachytherapy standards for reference air kerma rate of the NMIJ and the BIPM. *Metrologia* 2016; 53 Tech. Suppl. 06001.

#### **[LNE-LNHB \(France\)](#)**

Aubineau-Lanièce I, Chauvenet B, Cutarella D, Gouriou J, Plagnard J and Aviles Lucas P. LNE-LNHB air-kerma and absorbed dose to water primary standards for low dose-rate <sup>125</sup>I brachytherapy sources. *Metrologia* 2012; 49: S189–92.

Douysset G, Gouriou J and Delaunay F. Dose metrology for high dose rate brachytherapy: from the definition of the national standard towards transfer to users [In French]. *Revue Française de Métrologie* 2007; 2: 3-10, ISSN 1772-1792

Douysset G, Gouriou J, Delaunay F, DeWerd L, Stump K and Micka J. Comparison of dosimetric standards of USA and France for HDR brachytherapy. *Phys Med Biol* 2005; 50: 1961–78.

#### **[NIST \(USA\)](#)**

Seltzer S M, O’Brien M and Mitch M G. New national air-kerma standard for low-energy electronic brachytherapy sources. *J Res Natl Inst Stand Technol* 2014; 119: 554–74.

Seltzer S M, Lamperti P J, Loevinger R, Mitch M G, Weaver J T and Coursey B M. Erratum: New national air-kerma-strength standards for <sup>125</sup>I and <sup>103</sup>Pd brachytherapy seeds. *J Res Natl Inst Stand Technol* 2004; 109: 301.

Seltzer S M, Lamperti P J, Loevinger R, Mitch M G, Weaver J T and Coursey B M. New national air-kerma-strength standards for <sup>125</sup>I and <sup>103</sup>Pd brachytherapy seeds. *J Res Natl Inst Stand Technol* 2003; 108: 337–58.

Seltzer S M and Bergstrom P M. Changes in the U.S. primary standards for the air kerma from gamma-ray beams. *J Res Natl Inst Stand Technol* 2003; 108: 359–81.

### NPL (UK)

Sander T, Duane S, Lee N D, Thomas C G, Owen P J, Bailey M and Palmans H. NPL's new absorbed dose standard for the calibration of HDR  $^{192}\text{Ir}$  brachytherapy sources. *Metrologia* 2012; 49: S184–8.

Alvarez J T, Sander T, de Pooter J A, Allisy-Roberts P J and Kessler C. Comparison BIPM.RI(I)-K8 of high dose-rate Ir-192 brachytherapy standards for reference air kerma rate of the NPL and the BIPM. *Metrologia* 2014; 51 Tech. Suppl. 06024.

Sander T and Nutbrown R F. The NPL air kerma primary standard TH100C for high dose rate  $^{192}\text{Ir}$  brachytherapy sources. *NPL report RD004* 2006; National Physical Laboratory, Teddington, UK.

Shipley D R and Duane S. Corrections to air kerma rate measurements of  $^{125}\text{I}$  brachytherapy sources to free space conditions. *NPL report RSA(EXT) 51* 1994; National Physical Laboratory, Teddington, UK.

### NRC (Canada)

Shen H, Culberson W S and Ross C K. Technical Note: An investigation of polarity effects for wide-angle free-air chambers. *Med Phys* 2016; 43: 4106–12.

El Gamal I, Cojocar C, Mainegra-Hing E and McEwen M. The Fricke dosimeter as an absorbed dose to water primary standard for Ir-192 brachytherapy. *Phys Med Biol* 2015; 60: 4481–95.

Mainegra-Hing E and Rogers D W O. On the accuracy of techniques for obtaining the calibration coefficient  $N_K$  of  $^{192}\text{Ir}$  HDR brachytherapy sources. *Med Phys* 2006; 33: 3340–7.

Kessler C, Downton B and Mainegra-Hing E. Comparison BIPM.RI(I)-K8 of high dose-rate Ir-192 brachytherapy standards for reference air kerma rate of the NRC and the BIPM. *Metrologia* 2015; 52 Tech. Suppl. 06013.

### PTB (Germany)

Schneider T. The PTB primary standard for the absorbed-dose to water for I-125 interstitial brachytherapy sources. *Metrologia* 2012; 49: S198–202.

Schneider T and Selbach H-J. Realization of the absorbed-dose to water for I-125 interstitial brachytherapy sources. *Radiother Oncol* 2011; 100: 442–5.

Bambynek M and Krauss A. Determination of absorbed dose to water for  $^{192}\text{Ir}$  HDR brachytherapy sources in near-field geometry. In: Kapsch R-P, ed. *PTB report: advanced metrology for cancer therapy*. Proc Int Conf Braunschweig, 2011; PTB-Dos-56. Braunschweig, Germany: PTB; 2011.

Kessler C, Allisy-Roberts P J and Selbach H J. Comparison BIPM.RI(I)-K8 of high dose-rate Ir-192 brachytherapy standards for reference air kerma rate of the PTB and the BIPM. *Metrologia* 2015; 52 Tech. Suppl. 06005.

Selbach H-J, Kramer H-M and Culberson W S. Realization of reference air-kerma rate for low-energy photon sources. *Metrologia* 2008; 45: 422–8.

Selbach H-J. New calibration device for  $^{192}\text{Ir}$  and  $^{60}\text{Co}$ -brachytherapy radiation sources [In German]. In: Bogner L, Dobler B, ed. *Medizinische Physik* 2006. Tagungsband der 37. Jahrestagung der Deutschen Gesellschaft für Medizinische Physik e.V.; 20–23

Büermann L, Kramer H-M, Schrader H and Selbach H-J. Activity determination of  $^{192}\text{Ir}$  solid sources by ionization chamber measurements using calculated corrections of selfabsorption. *Nucl Instr Meth Phys Res A* 1994; 339: 369–76.

### [SSDL-ININ \(Mexico\)](#)

Álvarez Romero J T, Tovar Muñoz V M and de la Cruz Hernández D. The SSDL-ININ calibration and measurement capability (CMC) for reference air kerma rate  $K_R$  in HDR  $^{192}\text{Ir}$  for well chambers. *AIP Conference Proceedings* 2016; 1747, doi. 10.1063/1.4954093

Álvarez Romero J T, Hernández-Oviedo O and Tovar-Muñoz V M. Calibration of  $^{125}\text{I}$  seeds at Mexico City altitude. *ID A10029 IFMBE Proceedings* 2012

### [SSM \(Sweden\)](#)

Carlsson Tedgren Å and Grindborg J-E. Audit on source strength determination for HDR and PDR  $^{192}\text{Ir}$  brachytherapy in Sweden. *Radiother Oncol* 2008; 86: 126-30.

### [VSL \(The Netherlands\)](#)

Alvarez J T, de Pooter J A, Andersen C, Aalbers A H L, Allisy-Roberts P J and Kessler C. Comparison BIPM.RI(I)-K8 of high dose-rate Ir-192 brachytherapy standards for reference air kerma rate of the VSL and the BIPM. *Metrologia* 2014; 51 Tech. Suppl. 06022.

van Dijk E, Kolkman-Deurloo I-K, Damen P M G. Determination of the reference air kerma rate for  $^{192}\text{Ir}$  brachytherapy sources and the related uncertainty. *Med Phys* 2004; 31: 2826–33.

Petersen J J, van Dijk E and Aalbers A H L. Comparison of methods for derivation of  $^{192}\text{Ir}$  calibration factors for the NE 2561 & 2571 ionisation chambers. Report S-El-94.01. Utrecht, The Netherlands: NMI Van Swinden Laboratorium; 1994.