




Messunsicherheitsmodell und -bilanz bei der Kalibrierung von Durchflussmessgeräten

Berlin
19. März 2013

Rainer Engel
Leiter der Arbeitsgruppe "Rückführung Flüssigkeitsmessungen"
PTB Braunschweig

- Durchfluss-Normalmesseinrichtung
- Modell und Einflussgrößen
- Messunsicherheitsbilanz
- Beiträge der Einflussgrößen
- Dynamische Einflüsse





(Geräte-) K-Faktor (Impulszählung):

$$K_{Meter} = \frac{f_{Output}}{\dot{V}_{REF}}$$

Standardunsicherheit:

$$u_{K_meter}^2 = \left(\frac{\partial K_{Meter}}{\partial f_{Output}} u_f \right)^2 + \left(\frac{\partial K_{Meter}}{\partial \dot{V}} \cdot \frac{\partial \dot{V}}{\partial m} u_m \right)^2 + \left(\frac{\partial K_{Meter}}{\partial \dot{V}} \cdot \frac{\partial \dot{V}}{\partial \rho_{Water}} u_\rho \right)^2 + \left(\frac{\partial K_{Meter}}{\partial \dot{V}} \cdot \frac{\partial \dot{V}}{\partial (\Delta V_{IP})} u_{\Delta V} \right)^2 + \left(\frac{\partial K_{Meter}}{\partial \dot{V}} \cdot \frac{\partial \dot{V}}{\partial (\Delta V_{T_Error})} u_{T_Error} \right)^2 + \left(\frac{\partial K_{Meter}}{\partial \dot{V}} \cdot \frac{\partial \dot{V}}{\partial T_{MEAS}} u_T \right)^2$$

Relative Standardunsicherheit

$$\left(\frac{u_{K_Meter}}{K_{Meter}} \right)^2 = \left(\frac{u_f}{f_{Output}} \right)^2 + \left(\frac{u_m}{m} \right)^2 + \left(\frac{u_\rho}{\rho_{Water}} \right)^2 + \left(\frac{u_{\Delta V}}{V_0} \right)^2 + \left(\frac{u_{T_Error}}{V_0} \right)^2 + \left(\frac{u_T}{T_{MEAS}} \right)^2$$

**Dichtemessgerät:**

- Typ: DMA 500
(Anton Paar GmbH., Austria)
- Sensor: vibrating U-tube

Charakteristika:

- *Bereich:* 0...3 g/cm³
- *Genauigkeit:* - Dichte: 5·10⁻⁶ g/cm³
- Temperatur: 0.01 °C

Pumpe zur Probennahme:**Probe:** 1 ml**Probenbehälter**

Messung der Wasserdichte (Funktionsprinzip)

Eingangsgrößen der Dichte-Approximationsfunktion:

$$\rho_{Approx}(\mathcal{G}_{Water}) = f(\mathcal{G}_{Cal,i}, \rho_{Cal,i}, \mathcal{G}_{Water})$$

Unsicherheitsbeiträge:

- | | |
|--|---|
| - Temperaturanzeige des Dichtemessgerätes: | ± 10 mK, standard uncertainty
Type A |
| - Dichteanzeige: | 0.0085 kg/m ³ (Type B: rectangular
probability distribution) |
| - Numerische Approximation (3. Ordnung): | ± 2.0·10 ⁻⁶ or 1,997 kg/m ³
(rectangular probability function) |


Unsicherheitsbeitrag durch Anwendung der Approximationsfunktion:
(Sensitivitätskoeffizienten: numerische Berechnung)

$$[U(\rho_{Approx})]^2 = \left[\frac{\partial \rho_{Approx}}{\partial \vartheta_{Cal,i}} \cdot u(\vartheta_{Cal,i}) \right]^2 + \dots$$

$$\dots + \left[\frac{\partial \rho_{Approx}}{\partial \rho_{Cal,i}} \cdot u(\rho_{Cal,i}) \right]^2 + \left[\frac{\partial \rho_{Approx}}{\partial \vartheta_{Process}} \cdot u(\vartheta_{Process}) \right]^2$$

Unsicherheitsbeiträge (Einflussgrößen):

- Temperaturmessung: Pt-100-Sensor (mit Elektronik)
- Temperaturanzeige: DMA 5000, Dichtemessgerät
- Dichteanzeige: DMA 5000, Dichtemessgerät



- Einflüsse:**
- 1) – Zwischenrohr-Volumen
 - 2) – eingeschlossenes Luftvolumen


$$\Delta V_{IP} = \left[\frac{\pi}{4} \cdot \left[d_i + \frac{(p_{finish} - p_0) \cdot d_i^2}{2 \cdot w \cdot E} \right]^2 \cdot l \cdot [1 + 3 \cdot \alpha \cdot (\vartheta_{finish} - \vartheta_0)] - V_{air,0} \cdot \frac{p_0}{p_{finish}} \cdot \frac{\vartheta_{finish} + 273.15}{\vartheta_0 + 273.15} \right] \dots$$

$$\dots - \left[\frac{\pi}{4} \cdot \left[d_i + \frac{(p_{start} - p_0) \cdot d_i^2}{2 \cdot w \cdot E} \right]^2 \cdot l \cdot [1 + 3 \cdot \alpha \cdot (\vartheta_{start} - \vartheta_0)] - V_{air,0} \cdot \frac{p_0}{p_{start}} \cdot \frac{\vartheta_{start} + 273.15}{\vartheta_0 + 273.15} \right]$$

- 3) – eingeschlossenes Flüssigkeitsvolumen

CALCULATION OF UNCERTAINTY BUDGET

Test Rig: **3-tons balance**



Parameters:

Flowrate:	72,0	m ³ /h	
- Balance Target weight:	2500,0	kg	
- Cal. weights Density:	7845,0	kg/m ³	Carbon steel
Diversion time:	125,1	s	
Ref. Volume:	2,5030	m ³	
- Ambient air Air density:	1,1800	kg/m ³	
- Interc. piping IP volume:	0,5008	m ³	steel, plastics
"effective" TC:	1,7689E-04	1/K	
delta Temp.:	0,050	K	During measm.
- Diverter Timing error:	20,0	ms	Standard dev.
Fluid Water density:	998,8173	kg/m ³	(at 20 °C)
Water temp.:	20,000	°C	
Test meter Pulse frequenc:	209,00	Hz	Turbine meter
Repeatability:	3,50E-05		

Example

Measurement quantity:

- Mass

Symbol	Source of uncertainty	Value ±g	Probability distribution	Divisor	c _i	u _i (W _i) ±g	u _i (W _i) ±%
W _s	Calibration of standard weights	2,3	normal	2,0	1,0	1,150	4,60E-05
S ₀	Discrimination of balance	0,5	rectangular	1,73	1,0	0,289	1,16E-05
S _R	Repeatability of balance	104	normal	2,0	1,0	52,000	2,08E-03
S _{01h}	Drift of balance	10	rectangular	1,73	1,0	5,780	2,31E-04
u _c (W _i)	Combined uncertainty		normal			52,3	2,093E-03
U(W _i)	Expanded uncertainty		normal(k = 2)			104,7	0,004

- Buoyancy


Symbol	Source of uncertainty	Value ±kg/m ³	Probability distribution	Divisor	c _i	u _i (W _i) ±kg	u _i (W _i) ±%
D _{air}	Ambient air density	5,90E-04	normal	2,0	8,758E-04	2,584E-07	2,19E-05
D _{water}	Water density	3,31E-02	normal	2,0	1,920E-08	3,178E-10	3,18E-11
D _{cal}	Density of calibration weights	1,0	normal	2,0	1,185E-06	5,927E-07	7,56E-09
u _c (C _B)	Combined uncertainty		normal			6,466E-07	5,479E-08
U(C _B)	Expanded uncertainty		normal(k = 2)			0,0	0,000

- Temperature

Symbol	Source of uncertainty	Value ±K	Probability distribution	Divisor	c _i	u _i (T _x) ±K	u _i (T _x) ±%
T ₀	Discrimination of temp.measurement	0,025	rectangular	1,73	1,0	0,014	7,23E-02
T _C	Sensor/meter calibration	0,18	rectangular	1,73	1,0	0,104	5,20E-01
T ₀	Temp. gradient in water (estimated)	0,5	rectangular	1,73	1,0	0,289	1,45E+00
u _c (T _x)	Combined uncertainty		normal			0,308	1,538E+00
U(T _x)	Expanded uncertainty		normal(k = 2)			0,615	3,075

R. Engel – 268. PTB-Seminar – 19.03.2013 - 11 -

Fortsetzung:



- Density (Approximation function based on density measurement with type DMA 500 density meter)

Symbol	Source of uncertainty	Value ±	Probability distribution	Divisor	c _i	u _i (D _i) ±kg/m ³	u _i (D _i) ±%
D _{E, approx}	Temp. Meas. for appr. [K]	0,010	normal	2,0	6,6435E-06	3,3218E-08	3,33E-09
D _{D, approx}	Density meas. for appr. [kg/m ³]	0,0085	normal	2,0	1,9995	0,008	8,51E-04
D ₀	Temp. downstream of MUT [K]	0,308	normal	2,0	2,08E-01	0,032	3,20E-03
D _B	Numeric approx. Error [kg/m ³]	0,002	rectangular	1,73	1,0	0,001	1,16E-04
u _c (D _i)	Combined uncertainty		normal			0,033	3,315E-03
U(D _i)	Expanded uncertainty		normal(k = 2)			0,066	0,007

- Time (Measurement/Diversion)

Symbol	Source of uncertainty	Value ±ms	Probability distribution	Divisor	c _i	u _i (T _x) ±s	u _i (T _x) ±%
T _C	Calibration of Timer (5x10-9)	5,0E-12	normal	2,0	1,0	2,500E-15	2,00E-15
T _D	Discrimination of Timer display	0,001	rectangular	1,73	1,0	5,780E-07	4,62E-07
T _T	Discrimination of Diverter Time	0,05	rectangular	1,73	1,0	2,890E-05	2,31E-05
u _c (T _x)	Combined uncertainty		normal			2,891E-05	2,3099E-05
U(T _x)	Expanded uncertainty		normal(k = 2)			5,782E-05	4,620E-05

- Diverter timing error

Symbol	Source of uncertainty	Value ±	Probability distribution	Divisor	c _i	u _i (T _{D, Error}) ±m ³	u _i (T _{D, Error}) ±%
T _C	Diverter timing error [ms]	20,00	normal				
F	Flowrate [m ³ /h]	72,00	normal				
V _{Error}	Diversion error volume [m ³]	4,0000E-04	normal	2,0	1,0	2,000E-04	7,9905E-03
u _c (T _{Error})	Combined uncertainty		normal			2,000E-04	7,991E-03
U(T _{Error})	Expanded uncertainty		normal(k = 2)			4,000E-04	1,598E-02

- Interconnecting piping volume error

Symbol	Source of uncertainty	Value ±	Probability distribution	Divisor	c _i	u _i (V _{p, Error}) ±m ³	u _i (V _{p, Error}) ±%
delta T _{ip}	Temp. Variation during calibr. [K]	0,050	rectangular				
V _{ip, TC}	IP volume [m ³]	0,501	rectangular				
TC	effective TC [1/K]	1,7689E-04	rectangular				
V _{ip, error}	IP volume error [m ³]	4,4294E-06	rectangular	1,73	1,0	2,560E-06	1,023E-04
u _c (T _{Error})	Combined uncertainty		normal			2,560E-06	1,023E-04
U(T _{Error})	Expanded uncertainty		normal(k = 2)			5,121E-06	2,046E-04

- Meter under test

Symbol	Source of uncertainty	Value ±	Probability distribution	Divisor	c _i	u _i (f _x) ±Hz	u _i (f _x) ±%
f ₀ = 0,5 * 1/N _{period}	Discr. of frequency meas.	3,32E-05	rectangular	1,73	1,0	4,0141E-03	1,921E-03
u _c (f _x)	Combined uncertainty		normal			4,014E-03	1,921E-03
U(f _x)	Expanded uncertainty		normal(k = 2)			8,028E-03	0,004

R. Engel – 268. PTB-Seminar – 19.03.2013 - 12 -

Summarized Uncertainty - "Flowrate measurement"

Symbol	Source of uncertainty	Value ± %	Probability distribution	u_i (FLOW) ± %
$u_c(m)$	Combined uncertainty (MASS)	2,093E-03	normal	2,093E-03
$u_c(C_{Buoy})$	Combined uncertainty (BUOYANCY)	5,479E-08	normal	5,479E-08
$u_c(\rho)$	Combined uncertainty (DENSITY)	3,315E-03	normal	3,315E-03
$u_c(T_{MEAS})$	Combined uncertainty (TIME)	2,310E-05	normal	2,310E-05
$u_c(V_{IP})$	Combined uncertainty (IP volume)	1,023E-04	normal	1,023E-04
$u_c(T_{Error})$	Combined uncertainty (Time error)	7,991E-03	normal	7,991E-03
S(Repeatabil.)	Repeatability uncertainty	3,500E-03	normal	3,500E-03
$u_c(FLOW)$	Combined uncertainty (FLOW)		normal	0,00956
U(FLOW)	Expanded uncertainty (FLOW)		normal(k = 2)	0,019


Summarized Uncertainty - "Quantity passed measurement"

Symbol	Source of uncertainty	Value ± %	Probability distribution	u_i (FLOW) ± %
$u_c(m)$	Combined uncertainty (MASS)	2,093E-03	normal	2,093E-03
$u_c(C_{Buoy})$	Combined uncertainty (BUOYANCY)	5,479E-08	normal	5,479E-08
$u_c(\rho)$	Combined uncertainty (DENSITY)	3,315E-03	normal	3,315E-03
$u_c(T_{MEAS})$	Combined uncertainty (TIME)	2,310E-05	not effective	
$u_c(V_{IP})$	Combined uncertainty (IP volume)	1,023E-04	normal	1,023E-04
$u_c(T_{Error})$	Combined uncertainty (Time error)	7,991E-03	normal	7,991E-03
S(Repeatabil.)	Repeatability uncertainty	3,500E-03	normal	3,500E-03
$u_c(QUANTITY)$	Combined uncertainty (QUANTITY)		normal	0,00956
U(Quantity)	Expanded uncertainty (QUANTITY)		normal(k = 2)	0,019

Summarized Uncertainty - "Meter K-factor determination"

Symbol	Source of uncertainty	Value ± %	Probability distribution	u_i (K) ± %
$u_c(m)$	Combined uncertainty (MASS)	2,093E-03	normal	2,093E-03
$u_c(C_{Buoy})$	Combined uncertainty (BUOYANCY)	5,479E-08	normal	5,479E-08
$u_c(\rho)$	Combined uncertainty (DENSITY)	3,315E-03	normal	3,315E-03
$u_c(T_{MEAS})$	Combined uncertainty (TIME)	2,310E-05	normal	2,310E-05
$u_c(V_{IP})$	Combined uncertainty (IP volume)	1,023E-04	normal	1,023E-04
$u_c(T_{Error})$	Combined uncertainty (Time error)	7,991E-03	normal	7,991E-03
$u_c(Frequency)$	Combined uncertainty (Frequency)	1,921E-03	normal	1,921E-03
S(Repeatabil.)	Repeatability uncertainty	3,500E-03	normal	3,500E-03
$u_c(K-FACTOR)$	Combined uncertainty (K-FACTOR)		normal	0,00976
U(K-FACTOR)	Expanded uncertainty (K-FACTOR)		normal(k = 2)	0,020

Erweiterte Messunsicherheit



Unsicherheitsbeiträge relevanter Funktionskomponenten

Sources of uncertainty	Relative standard uncertainty [%]	
Mass		2,09E-03
Calibration of standard weights	4,60E-05	
Discrimination of balance	1,16E-05	
Repeatability of balance	2,08E-03	
Drift of balance	2,31E-04	
Bouyancy		2,19E-05
Ambient air density	2,19E-05	
Water density	3,18E-11	
Density of calibration weights	7,56E-09	
Temperature of fluid	1,54E+00	
Discrimination of temp.meas.	7,23E-02	
Sensor/meter calibration	5,20E-01	
Temp. gradient in water (estim.)	1,45E+00	
Density of fluid		3,31E-03
Temp. Meas. for appr. function	3,33E-09	
Density meas. for appr. function	8,51E-04	
Temp. downstream of MUT	1,54E+00	
Numeric approx. Error	1,16E-04	
Time (Measurement/Diversion)		2,31E-05
Calibration of Timer (5x10-9)	2,00E-15	
Discrimination of Timer display	4,62E-07	
Discrimination of Diverter Time	2,31E-05	
Diverter timing error		7,99E-03
Diverter timing error		
Flowrate		
Diversion error volume	7,99E-03	
Interconnecting piping volume		1,02E-04
Temp. Variation during calibr.	1,54E+00	
IP volume		
effektive TC		
IP volume error	1,02E-04	
Meter under test (MUT)		1,92E-04
Discr. of frequency meas.	1,92E-04	
Repeatability uncertainty		3,50E-03
u_c(K-factor)	Combined uncertainty	0,010
U(K-factor)	Expanded Uncertainty (k=2)	0,019

$u(x_1)$

$u(x_2)$

$u(x_3)$

$u(x_4)$

$u(x_5)$

$u(x_6)$


$u(x_7)$


$u(x_8)$

$u(x_9)$


Korrelations-
effekte

$u(y)$







Element-by-element traceability: Dynamic effects




Meter errors caused by flow rate variations:
linear steady-state characteristics



Meter errors caused by flow rate variations:
due to nonlinear steady-state characteristics



PTB Komponentenweise Rückführkette: dynamische Effekte



R. Engel – 268. PTB-Seminar – 19.03.2013 - 21 -

PTB Unsicherheitsbudget – Erweiterung

Sources of uncertainty	Relative standard uncertainty [%]
Mass	2,09E-03
Calibration of standard weights	4,60E-05
Discrimination of balance	1,16E-05
Repeatability of balance	2,08E-03
Drift of balance	2,31E-04
Dynamic impacts	
Buoyancy	2,19E-05
Ambient air density	2,19E-05
Water density	3,18E-11
Density of calibration weights	7,56E-09
Temperature of fluid	1,54E+00
Discrimination of temp. meas.	7,23E-03
Sensor/meter calibration	5,20E-01
Temp. gradient in water (resin)	1,45E+00
Dynamic impacts	
Density of fluid	3,31E+03
Temp. Meas. for appr. function	3,33E-09
Density meas. for appr. function	8,56E-04
Temp. downstream of MUT	1,54E+00
Numeric approx. Error	1,16E-04
Dynamic impacts	
Time (Measurement/Diversion)	2,31E-05
Calibration of Timer (S&D-S)	2,00E-15
Discrimination of Timer display	4,62E-07
Discrimination of Diverter Time	2,31E-05
Diverter timing error	7,99E-03
Diverter timing error	
Flowrate	
Diversion error volume	7,99E-03
Dynamic impacts	
Interconnecting piping volume	1,02E-04
Temp. Variation during calibr.	1,54E+00
IP volume	
effective TC	
IP volume error	1,02E-04
Meter under test (MUT)	1,92E-04
Disc. of frequency meas.	1,92E-04
Repeatability of MUT	
Reproducibility of MUT	
Dynamic impacts (pressure)	
Repeatability uncertainty of facility	3,02E-03
u_c (K-factor) Combined uncertainty	0,010
U (K-factor) Expanded Uncertainty (k=2)	0,019

Nur statische Einflüsse

Statische & dynamische Einflüsse

R. Engel – 268. PTB-Seminar – 19.03.2013 - 22 -

Vielen Dank
für Ihre
Aufmerksamkeit.

Dr. Rainer Engel
Leiter der Arbeitsgruppe 1.53 "Rückführung Flüssigkeitsmessungen"
PTB Braunschweig
rainer.engel@ptb.de