

Marktüberwachung in Europa: Aktuelle Entwicklung

Änderungen der maßgeblichen RL und deren
Konsequenz

Aktuelle Änderung in der Rechtssetzung

- **765/2008/EC und Beschluss 768/2008/EC**
- **MID und NAWI “fahren Omnibus”**
- **Schutzklauselverfahren neu geregelt**

Schutzklauselverfahren NAWI

Artikel 8

- **MS trifft geeignete Maßnahme**
- **MS unterrichtet KOM**
- **KOM konsultiert Betroffene**
- **KOM unterrichtet MS über Ergebnis der Konsultation**
- **Hält KOM Maßnahme für gerechtfertigt, Info an MS**
- **Auch Info an alle anderen MS**

Schutzklauselverfahren

3.6. Marktüberwachung und Schutzklauselverfahren

In dem Vorschlag wird das bestehende Schutzklauselverfahren verbessert. Es wird eine Stufe des Informationsaustauschs zwischen den Mitgliedstaaten eingeführt und dargelegt, welche Schritte die betreffenden Behörden unternehmen müssen, wenn ein nichtkonformes Messgerät erkannt wird. Ein echtes Schutzklauselverfahren (das dazu führt, dass die Kommission darüber entscheidet, ob eine Maßnahme begründet ist oder nicht) wird nur dann eingeleitet, wenn ein Mitgliedstaat einen Einwand gegen eine Maßnahme erhebt, die ein anderer Mitgliedstaat gegen ein Messgerät ergriffen hat. Besteht Einigkeit hinsichtlich der beschränkenden Maßnahme, die von einem Mitgliedstaat ergriffen wurde, müssen alle Mitgliedstaaten auf ihrem Hoheitsgebiet entsprechend tätig werden.

Ablaufschema bei Maßnahmen der MÜ

- Was?
 - Wann?
 - Wer?
 - Wem?

Investigation – Sachverhalt ermitteln

1

The MSA finds an instrument suspicious not to comply with the directive and investigates the matter possibly in participation with the economic operator.

Hearing - Offizielle Anhörung

2

The Market Surveillance Authority gives the opportunity to be heard to the economic operator concerned. E.g. sending a letter to the economic operator with the details found and asking for his opinion within 14 days (Regulation 756/2008/EC; Article 21 (3)).

If the opinion of the economic operator will not dispel suspicions or if the operator does not answer within the given time the MSA will continue in the process.

Corrective measures - Maßnahmen

MSA gives the opportunity to the economic operator to take corrective measures:

- a. The MSA asks the economic operator to take corrective measures. This is not necessary if the operator is taking corrective actions by himself which could prevent or redress the non-compliance.
- b. The MSA informs the economic operator of the measure and the remedies available; usually done by no. 2 and 3a.
- c. The MSA informs the concerned NoBo about the non-compliance.
- d. The MSA informs other Member States which might have the same measuring instrument.
- e. The MSA informs the notifying authority of the involved NoBo.
- f. The MSA informs the Commission.

3

Corrective measures - Maßnahmen

4

MSA takes actions when no corrective or ineffective corrective measures (under no. 4) have been taken by the economic operator:

- a. The MSA takes appropriate measures (Article 19 (2) 765/2008).
- b. The MSA informs the economic operator of the measure and the remedies available; usually done by no. 2 and 4a.
- c. The MSA informs the other MS.
- d. The MSA informs the notifying authority of the involved NoBo.
- e. The MSA informs the Commission.

5

The MSA informs all NoBo or NoBoMet about the non-compliance when it is a valuable information for their future work.

This information must be anonymous.

6

The MSA informs public if it is affected by that non-compliance.

note

All information for a group (MS or NoBos) and public information should be run by ICSMS (at least in the future).

Some (information) steps are taken at the same time.



Soooo muss Marktüberwachung

Results of market surveillance

Dr. Eberhard Pettit



Example II: rental scale




- found by acting as an MSA
- no proper sealing
- wrong understanding of the manufacturer (modul D)
- criminal?

Dr. Pettit, Germany, 23 October 2008/Malta



Example VI: road (oil) tankers



Dr. Pettit, Germany, 23 October 2008/Malta

Example V: taxameter



- Secured?
- Reed contacts or infrared-sensors e.g. are not suitable to prevent unauthorized access to the adjustment mode

Dr. Pettit, Germany, 23 October 2008/Malta

