

EMRP JRP IND62 – TIM: Use of on-board metrology systems for area-scanning on machine tools



Author: Co-authors: doc. Ing. Vít Zelený, CSc. doc. Ing. Ivana Linkeová, Ph.D. Ing. Jakub Sýkora Pavel Skalník Jaromír Hynek






- CNC MACHINE TOOL
- FREEFORM STANDARD
- SCANING
- ON-BOARD METROLOGY
- INTERCOMPARISON
- E_n CRITERION






• CNC MACHINE TOOL


- FREEFORM STANDARD
- SCANING
- ON-BOARD METROLOGY












- **CNC MACHINE TOOL**
- FREEFORM STANDARD •
- **ON-BOARD METROLOGY**

CMI Hyperbolic paraboloid freeform standard










- CNC MACHINE TOOL
- FREEFORM STANDARD
- SCANING
- ON-BOARD METROLOGY



CONTACT SCANNING (e.g. ZEISS PRISMO with active scanning probe VAST)







- CNC MACHINE TOOL
- FREEFORM STANDARD
- SCANING
- ON-BOARD METROLOGY

LASER SCANNING

(e.g. based on autofocus)



May 18, 2016





- CNC MACHINE TOOL
- FREEFORM STANDARD
- SCANING
- ON-BOARD METROLOGY




LASER SCANNING

(e.g. based on autofocus)



- CNC MACHINE TOOL
- FREEFORM STANDARD
- SCANING
- ON-BOARD METROLOGY

AREA SCANNING – PHOTOGRAMMETRY



VDI/VDE 2634 Part 2 -Optical 3-D measuring systems - Optical systems based on area scanning


VDI/VDE 2634 Part 3 -VDI/VDE 2634 Part 3 Optical 3D-measuring systems -Multiple view systems based on area scanning








Fringe projection





- CNC MACHINE TOOL
- FREEFORM STANDARD
- SCANING
- ON-BOARD METROLOGY







- CNC MACHINE TOOL
- FREEFORM STANDARD
- SCANING
- ON-BOARD METROLOGY








Demonstration – measurement of FF-MS on machine tool Deckel Maho DMU 50 with Renishaw probe OMP 400

Productive Process Pyramid[™]

(acc. Metrology solutions for productive process control, Brochure of Renishaw)



In-process control is represented by active controls applied during metal cutting. The control is focused on inherent sources of errors caused during all machining processes, such as tool wear, part deflection and the impact of temperature and heat flows. The inprocess control is carried out by onmachine probing module equipped by inspection cycles included in NC programs.





- WHY FREEFORM?
- WHY FREEFORM STANDARD?
- HOW TO USE FREEFORM STANDARD?
- INTERCOMPARISON
- E_n CRITERION


Basic geometrical elements

Plane




• WHY FREEFORM?


- WHY FREEFORM STANDARD?
- HOW TO USE
 FREEFORM
 STANDARD?



$$ax + by + cz + d = 0$$

Gauge block Step gauge






Sphere

 $x^2 + y^2 = r^2$

Hole bar Hole plate



х

Ball Ball bar Ball plate


Freeform metrology – what is a freeform geometry?






Cylinder as a basic


geometrical element



Cylinder as a freeform geometry (represented in NURBS)







 $x^2 + y^2 = r^2$


 $\mathbf{S}(u,v) = \sum_{i=0}^{m} \sum_{j=0}^{n} R_{i,j}(u,v) \mathbf{P}_{i,j}, \ 0 \le u \le 1, \ 0 \le v \le 1$ $R_{i,j}(u,v) = \frac{N_{i,p}(u)N_{j,q}(v)w_i}{\sum_{i=0}^{m} \sum_{j=0}^{n} N_{i,p}(u)N_{j,q}(v)w_i}$

15

May 18, 2016

Freeform metrology – what is a freeform geometry?










FREE-FORM SURFACE

PARAMETRIC (u,v) AND CARTESIAN (x,y,z) COORDINATES







18


Hyperbolic paraboloid

Mathematical model


Explicit equation

z(x, y) = p + k(x - m)(y - n),where (m, n, p) are Cartesan coordinates of vertex and k is a shape coefficient


CAD model



 $N_{i,1}(u)$ and $N_{i,1}(v)$ are linear B-spline basis functions



P3 20 z -40P2 -20



May 18, 2016






FF-MS Hyperbolic paraboloid

in procedures for ensuring reliable measurements on machine tools, demonstration and integration focused on end-user needs

HOW TO USE FREEFORM STANDARD?



En criterion in in-process measurement



FREEFORM MEASUREMENT UNCERTAINTY EVALUATION



$$E_n = \frac{|X_R - X_L|}{\sqrt{U_R^2 + U_L^2}}, E_n \le 1$$

- X_R reference value measured on CMM
- *X_L* measured value on a tested CNC machine
- U_R reference measurement uncertainty of CMM
- *U_L* measurement uncertainty of a tested CNC machine

U_L ... UNKNOWN VALUE

$$U_L = \pm \sqrt{\frac{(X_R - X_L)^2}{E_n^2} - U_R^2}, \qquad \frac{(X_R - X_L)^2}{E_n^2} \ge U_R^2.$$





PILOT MEASUREMENT

SIP CMM 5




PILOT LABORATORY SPECIFICATION

| Measurement system | Tactile |
|----------------------------|-----------------|
| Max. perm. error | (0.5 + 0.8L) μm |
| Measurement uncertainty | 1.6 µm |
| Maximal | Length 720 mm |
| measurement | Width 720 mm |
| dimensions | Heigth: 550 mm |



Measurement of 3D line on machine tool






Reliable measurements on machine tools

| EURAMET | |
|---------|--|
|---------|--|

| MEASURE | MET LAB 1.1 |
|---|---|
| Measured standard | HP PA |
| Output data format | TXT Points along curves and lines on freeform surface |
| Evaluation SW | Geomagic |
| Comparison method | Bestfit alignment against the CAD model |
| Measurand definition | Deviation between the actual and nominal data, form error |
| Results | Colour map of |
| representation | deviations |
| Form error | [-8.1, 6.8] µm |
| Average deviation (one-sigma limits) | 2.3, -2.2 µm |
| Standard deviation | 2.8 µm |

HP PA










| MEASUREMET LAB 1.1 | |
|--------------------------|---|
| Measurand X _L | Deviation of each measured point obtained by Measurement 1.1 from reference CAD model |
| Limits of evaluation | Two-sigma limits |
| U _R | ± 1.6 μm |



HP PA


```
U_L = \pm 3.7 \,\mu\mathrm{m}
```





Coordinate system determination

Curves on freeform surface measurement







| MEASURE | MET LAB 1.2 | | |
|---|---|--------|--------------------------------------|
| Measured standard | HP TIM | HP TIM | - |
| Output data format | TXT Points along curves and lines on freeform surface | | |
| Evaluation SW | Geomagic | | |
| Comparison method | Bestfit alignment against the CAD model | 10 | 0.0050 |
| Measurand definition | Deviation between the actual and nominal data, form error | | 0.0028 0.0020 0.0012 0.0005 |
| Results representation | Colour map of deviations | | -0.0005 |
| Form error | [-24.7, 6.6] µm | | -0.0020 |
| Average deviation (one-sigma limits) | 1.6, -1.8 | 9/1 | -0.0028 -0.0035 -0.0043 |
| Standard deviation | 2.8 µm | | -0.0050 |


May 18, 2016

Final workshop - PTB, Braunschweig, Germany






| MEASUREMET LAB 2.2 | |
|--------------------------|---|
| Measurand X _L | Deviation of each measured point obtained by Measurement 1.2 from reference CAD model |
| Limits of evaluation | Two-sigma limits |
| U _R | ± 1.6 µm |



HP TIM

$$U_L = \pm 1.5 \,\mu\mathrm{m}$$







LABORATORY 2

ATOS Triple Scan



| LABORATORY 2 SPECIFICATION | |
|----------------------------|---------------------------|
| Measurement system | 3D optical |
| Maximal measurement | 170 x 130 - 810 x 610 mm² |
| dimensions | 170 × 130 - 810 × 010 mm |


EURAMET

MEASUREMET LAB 2

| Measured standard | HP PA |
|---|---|
| Output data format | STL – 3D point cloud |
| Evaluation SW | Geomagic |
| Comparison method | Bestfit alignment against the pilot data |
| Measurand definition | Deviation between the actual and pilot data |
| Results representation | Colour map of deviations |
| Form error | [-6.5 <i>,</i> 8.4] μm |
| Average deviation (one-sigma limits) | 1.2 μm, -1.3 μm |
| Standard deviation | 1.6 μm |


HP PA
















| MEASUREMET LAB 2 | |
|--------------------------|---|
| Measurand X _L | Deviation of point corresponding to SIP point obtained by Measurement 2 from reference CAD model |
| Limits of evaluation | Two-sigma limits |
| U _R | ± 1.6 μm |



HP PA

 $U_L = \pm 2.8 \,\mu\mathrm{m}$







LABORATORY 3 WERTH TOMOSCOPE HV 500



LABORATORY 3 SPECIFICATION

| Measurement system | Multisensor + tactile scanning probe |
|--------------------------------------|--------------------------------------|
| Max. perm. error | (4.5 + L/75) μm |
| Maximum X-ray voltage | 300 kV |
| Maximum X-ray performance | 300 W |
| Resolution | (2048 × 2048) pix. |
| Sensor area | (400 × 400) mm |
| Maximal measurement dimensions | Length 710 mm Diameter 500 mm |
| Maximal workpiece weight | 75 kg |


MEASUREMET LAB 3

| Measured standard | HP PA |
|--------------------------------------|---|
| Output data format | STL – 3D point cloud |
| Evaluation SW | Geomagic |
| Comparison method | Bestfit alignment against the pilot data |
| Measurand definition | Deviation between the actual and pilot data |
| Results | Colour map of |
| representation | deviations |
| Form error | [-7.6, 12.4] µm |
| Average deviation (one-sigma limits) | 1.8 μm, -1.6 μm |
| Standard deviation | 2.2 µm |






EURAMET







| MEASUREMET LAB 3 | | |
|--------------------------|--|--|
| Measurand X _L | Deviation of each calibrated point obtained by pilot tactile measurement from 3D point cloud obtained by Measurement 3 | |
| Limits of evaluation | Two-sigma limits | |
| U _R | ± 1.6 μm | |




HP PA

 $U_L = \pm 4.1 \,\mu {\rm m}$





CALIBRATED FF-MS, CALIBRATED CAD





Final workshop - PTB, Braunschweig, Germany

EURAME'



LABORATORY 5 ZEISS PRISMO





LABORATORY 5 SPECIFICATION

| Measurement system | Active scanning probe VAST |
|-----------------------|----------------------------|
| Max. perm. error | (1.0+L/330) µm |
| Maximal | Length 1200 mm |
| measurement | Width 850 mm |
| dimensions | Height 700 mm |


Final workshop - PTB, Braunschweig, Germany

EURAMET

Reliable measurements on machine tools




| MEASUREMET LAB 5 | | | | |
|---|---|--|--|--|
| Measured standard | HP TIM | | | |
| Output data format | TXT Points obtained by scanning probe | | | |
| Evaluation SW | Geomagic | | | |
| Comparison method | Bestfit alignment against the calibrated CAD model | | | |
| Measurand definition | Deviation between the actual and nominal data, form error | | | |
| Results | Colour map of | | | |
| representation | deviations | | | |
| Form error | [-7.6, 2.6] µm | | | |
| Average deviation (one-sigma limits) | ± 0.4 μm | | | |
| Standard deviation | 0.5 µm | | | |







| MEASUREMET LAB 5 | | | | | |
|--------------------------|--|--|--|--|--|
| Measurand X _L | Deviation of each measured point obtained by Measurement 5 from reference calibrated CAD model | | | | |
| Limits of evaluation | Two-sigma limits | | | | |
| U _R | ± 1.6 μm | | | | |



HP TIM

 $U_L = \pm 1.0 \ \mu \mathrm{m}$









| | Meas. technology | Measuring machine | Measurand | Measuremen uncertainty |
|-------|-------------------------------------|----------------------|---|--|
| PILOT | Tactile on CMM | SIP | Deviation of surface point from reference model | U _L = ± 1.6 μm |
| LAB 1 | Tactile on CNC | Resnishaw | Deviations of measured points from CAD | U _L = ± 3.7 μm U _L = ± 1.5 μm |
| LAB 2 | Optical 3D scanner | ATOS Triple Scan | Deviation of pilot data from STL | U _L = ± 2.8 μm |
| LAB 3 | Laser scanner | FARO Arm | Deviation of pilot data from STL | $U_L = \pm 40 \ \mu m$ |
| LAB 4 | Computer tomography | Werth CT | Deviation of pilot data from STL | $U_L = \pm 4.1 \ \mu m$ |
| LAB 5 | Tactile active scanning probe | ZEISS | Deviations of measured points from calibrated CAD | U _L = ± 1.0 μm |

May 18, 2016



THANKS FOR YOUR ATTENTION



Okružní 31, 638 00 Brno, Czech Republic, + 420 545 555 337, <u>www.cmi.cz</u>